

RNA # 10558



CERTIFICATE OF CALIBRATION

Standard Calibration, CardioCoach CO2 Device

3487 W 2100 S, Suite 300
Salt Lake City, UT 84120
TEL: 801/483-2080
FAX: 801/483-2123

This Certifies that the instrument listed herein was calibrated by KORR Medical Technologies, Inc. to designated Standards, according to specific KORR procedures. Equipment used to perform this calibration are certified by third-party calibration laboratories, or are natural physical constants. All calibrations are performed to original KORR specifications, unless otherwise noted.

MODEL	SERIAL NUMBER	CALIBRATION DATE
Cardio Coach CO2	15108	12/17/17

STANDARDS USED FOR CALIBRATION

MANUFACTURER	MODEL	SERIAL NUMBER	CALIBRATION DUE
Hans Rudolph, Inc.	1 Liter Calibration Syringe	1052	10/2018
Hans Rudolph, Inc.	3 Liter Calibration Syringe	1058	10/2018
Davis Instruments	Perception II Indoor Climate Monitor	1049	12/2017

ATMOSPHERIC MEASUREMENTS

CALIBRATION	STANDARD	TOLERANCE	AS FOUND	AFTER CALIBRATION
Temperature °C	21°C	+/- 1°C		21°C
Barometric Pressure mmHg	658	+/- 3 mmHg		658
Humidity %	12%	+/- 5 %		12%
Atmospheric Oxygen	20.75%	+/- 0.15 %		20.8%
Nitrogen Cal Gas	0% O2	0.0% - 0.2%		0.08%
Atmospheric CO2 %	0.00%	+/- .05 %		0.00%
Cal Gas CO2 %	4.00%	+/- .10 %		4.01%

FLOW MEASUREMENTS

FLOW RATE SETTING	SYRINGE SIZE	RANGE	AS FOUND	AFTER CALIBRATION
10 LPM	1000 ml	980ml - 1020ml		1005 ml
30 LPM	1000 ml	980ml - 1020ml		1001 ml
70 LPM	3000 ml	2940ml - 3060ml		3000 ml
100 LPM	3000 ml	2940ml - 3060ml		3005 ml

New Device

CALIBRATED BY

TITLE

DATE

Mike Mathews	Service Tech	12/17/17
--------------	--------------	----------

VERIFIED BY

TITLE

DATE

[Signature]	Engineer	12/19/17
-------------	----------	----------